
V1.1 Revised 07/06/2020

# ELISA Assay Diluent (Non-Mammalian)

Provides a complex matrix for equalization of sample and standard matrices.

ELISA Assay Diluent is a non-mammalian proteinbased reagent formulated to equalize complex matrix differentials that are often encountered when evaluating serum and plasma samples in antibody-sandwich ELISA formats. Large differences between the sample and the standard diluent matrices will



every well to target antigen binds to equalize the and interfering the targe matrix molecules that antigen mix with the assay diluent

ated proteins and bdy small molecules to in the sample rget may suppress antigen binding efficiency

result in under-recovery of the target analyte present in sample wells. Complex and concentrated protein environments (the matrix) present in serum or plasma samples will greatly reduce the antigenbinding efficiency of the plate-adsorbed antibodies, resulting in a gross underestimation of the amount of target analyte present in the test samples. Assay Diluent minimizes this serious issue in ELISAformat assays testing serum or plasma samples.

To use, simply add 50-100 μL to every well of the ELISA plate, including all wells designated for standards, controls, and samples. Then add the standards, controls, and samples to the plate. The diverse protein composition of ELISA Assay Diluent will minimize nonspecific binding issues, decreasing background signal and increasing assay sensitivity. ELISA Assay Diluent also inhibits complement and thrombin activity present in serum and plasma samples. Incorporation of an antimicrobial agent allows for room temperature bench-top use and extensive storage stability at 2-8°C. This novel assay diluent was designed to ensure more accurate and consistent ELISA performance, regardless of sample type.

## ELISA ASSAY DILUENT (Non-Mammalian)

Catalog #
LS-M21-100
LS-M21-500
LS-M21-1

### **INSTRUCTIONS:**

- 1. Dilute the standard curve, controls, and the samples as necessary.
- 2. Pipette 50-100  $\mu$ L Assay Diluent per well into every well of the plate.
- 3. Pipette 50-200 μL of each standard, control, and sample into the plate.
- 4. Run the rest of the assay according to the specific ELISA protocol.
- 5. Analyze the data. Because all the wells, including the standards and controls, received the same volume of assay diluent, there is no need to account for this dilution when calculating the results.

For more ELISA protocols and information, please visit www.LSBio.com.

#### **SPECIFICATIONS:**

- Light yellow liquid
- 1X ready to use
- PH 7.2-7.6

#### **STORAGE:**

- 24 months at 2-8°C
- 1 week at room temperature

#### **SAFETY & USAGE:**

- Contains ≤ 0.1% sodium azide
- SDS available at LSBio.com
- Product intended for research use or for further manufacturing into in vitro diagnostics reagents only.
- Not intended for use in human or therapeutics purposes.

LifeSpan BioSciences, Inc. 2401 4th Avenue, Suite 900, Seattle, WA 98121

USA TOLL-FREE 1-866-819-4732 LOCAL 206-374-1102

www.LSBio.com