

Alcohol Dehydrogenase (ADH) Assay Kit (Colorimetric)

LS-K192-100 (100 Tests) • Store at -20°C



Introduction

ALCOHOL DEHYDROGENASE (ADH) is an oxidoreductase which catalyzes the interconversion of alcohols and aldehydes or ketones. ADH is important in humans and other organisms for the breakdown of alcohols which may otherwise be toxic. In yeast and some bacteria, ADHs catalyze the opposite reaction and produce alcohol as part of fermentation. This non-radioactive, colorimetric ADH assay is based on the reduction of the tetrazolium salt MTT in a NADH-coupled enzymatic reaction to a reduced form of MTT which exhibits an absorption maximum at 565 nm. The increase in absorbance at 565 nm is directly proportional to the enzyme activity.

Key Features

- Fast and sensitive. Linear detection range (20 µL sample): 0.4 to 80 U/L for 30 min reaction. Detection Limit of 0.1 U/L for 120 min reaction.
- Convenient and high-throughput. Homogeneous "mix-incubate-measure" type assay. Can be readily automated on HTS liquid handling systems for processing thousands of samples per day.

Applications

- ADH activity determination in biological samples (e.g. plasma, serum, urine, tissue and culture media.)

Components

Component	K192-100
	100 Tests
Assay Buffer	10 mL
Diaphorase	120 µL
NAD Solution	1 mL
Calibrator	1.5 mL
MTT Solution	1.5 mL
Substrate (10% Ethanol)	1 mL

Materials Not Supplied

Pipetting devices and accessories (e.g. multi-channel pipettor), clear flat-bottom 96-well plates (e.g. Corning Costar), centrifuge tubes and plate reader.

Storage

The kit is shipped at room temperature. Store all components at -20°C upon receiving. Shelf life: 6 months after receipt.

FOR RESEARCH USE ONLY! Not for use in humans.

LifeSpan BioSciences, Inc. • 2401 Fourth Avenue, Suite 900, Seattle, WA 98121
www.LSBio.com • (206) 464-1554 • TechnicalSupport@LSBio.com

Alcohol Dehydrogenase (ADH) Assay Kit (Colorimetric)



LS-K192-100 (100 Tests) • Store at -20°C

Assay Procedure

This assay is based on a kinetic reaction. To ensure identical incubation time, addition of Working Reagent to samples should be quick and mixing should be brief but thorough. Use of a multi-channel pipettor is recommended. Assays can be executed at any desired temperature (e.g. 25°C or 37°C).

Sample Preparation

Serum and plasma are assayed directly.

Tissue: prior to dissection, rinse tissue in phosphate buffered saline (pH 7.4) to remove blood. Homogenize tissue (50 mg) in ~200 µL buffer containing 50 mM potassium phosphate (pH 7.5). Centrifuge at 10,000 x g for 15 min at 4°C. Remove supernatant for assay.

Cell Lysate: collect cells by centrifugation at 2,000 x g for 5 min at 4°C. For adherent cells, do not harvest cells using proteolytic enzymes; rather use a rubber policeman. Homogenize or sonicate cells in an appropriate volume of cold buffer containing 50 mM potassium phosphate (pH 7.5). Centrifuge at 10,000 x g for 15 min at 4°C. Remove supernatant for assay.

All samples can be stored at -20 to -80°C for at least one month.

Reagent Preparation

Equilibrate reagents to desired reaction temperature (e.g. 25°C or 37°C). Briefly centrifuge tubes before use.

The Working Reagent (WR) is prepared by mixing, for each 96-well assay, 5 µL Substrate, 14 µL MTT Solution, 9 µL NAD Solution, 1 µL Diaphorase and 55 µL Assay Buffer.

The Blank Working Reagent (BWR) is prepared by mixing, for each 96-well assay, 14 µL MTT Solution, 9 µL NAD Solution, 1 µL Diaphorase and 60 µL Assay Buffer (i.e. no Substrate). Fresh reconstitution of the WRs is recommended.

Procedure

1. Transfer 100 µL H₂O (OD_{H₂O}) and 100 µL Calibrator (OD_{CAL}) solution into wells of a clear flat bottom 96-well plate.
2. Transfer 20 µL sample into 2 separate wells. Add 80 µL WR to one sample well and 80 µL BWR to the other sample well. Tap plate briefly to mix.
3. Read OD_{565nm} (OD₀), and again after 30 min (OD₃₀) on a plate reader.

Calculations

Subtract the OD₀ from OD₃₀ for each sample and sample blank well to compute the ΔOD_S and ΔOD_B values respectively. ADH activity can then be calculated as follows:

$$\begin{aligned} \text{ADH Activity} &= \frac{\Delta\text{OD}_S - \Delta\text{OD}_B}{\epsilon_{\text{mtt}} \cdot l} \times \frac{\text{Reaction Vol } (\mu\text{L})}{t \text{ (min)} \cdot \text{Sample Vol } (\mu\text{L})} \times n \\ &= \frac{273}{t \text{ (min)}} \times \frac{\Delta\text{OD}_S - \Delta\text{OD}_B}{\text{OD}_{\text{CAL}} - \text{OD}_{\text{H}_2\text{O}}} \times n \quad (\text{U/L}) \end{aligned}$$

where ϵ_{mtt} is the molar absorption coefficient of reduced MTT. l is the light path length which is calculated from the calibrator. OD_{CAL} and OD_{H₂O} are OD_{565nm} (OD₀) values of the Calibrator and water. t is the reaction time (30 min is the recommended time). Reaction Vol and Sample Vol are 100 µL and 20 µL, respectively. n is the dilution factor.

Unit definition: 1 Unit (U) of ADH will catalyze the conversion of 1 µmole of ethanol to acetaldehyde per min at pH 8.2.

Note: If sample ADH activity exceeds 80 U/L, either use a shorter reaction time or dilute samples in water and repeat the assay. For samples with ADH activity < 1 U/L, the incubation time can be extended to 2 hours.

FOR RESEARCH USE ONLY! Not for use in humans.

LifeSpan BioSciences, Inc. • 2401 Fourth Avenue, Suite 900, Seattle, WA 98121


www.LSBio.com • (206) 464-1554 • TechnicalSupport@LSBio.com

Alcohol Dehydrogenase (ADH) Assay Kit (Colorimetric)

LS-K192-100 (100 Tests) • Store at -20°C



Sample Data



Version: V.08.09.2018

FOR RESEARCH USE ONLY! Not for use in humans.

LifeSpan BioSciences, Inc. • 2401 Fourth Avenue, Suite 900, Seattle, WA 98121

www.LSBio.com • (206) 464-1554 • TechnicalSupport@LSBio.com